

Allweighs Weighing Services
 Westend Grange
 West End Road
 Epworth
 Doncaster
 DN9 1LA
 Tel: 01427 875 867
 Web: www.allweighs.co.uk



Certificate No: KCS:

202401046

Location: Site Weighbridge in

Manufacturer: Dini Argeo DFWL

Serial/Plant No: 0100911602

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Test Date: 21 August 2024

Next due: August 2025

Calibration Certificate of Accuracy

Business Name: Transwaste Ltd

Address: Melton Industrial Estate
 Gibson Lane South
 Melton
 Hull
 HU14 3HH

Contact Name: Graham Wildridge

Off-centre Loading:

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	6700	0	5	6700	
2	6700	0	6	6700	
3	6700	0			
4	6700	0			

Four loadcell

Six loadcell

Single loadcell

Eight loadcell

As Found

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
20700	20700	0
30700	30700	0
44000	44000	0

Repeatability

44000	44000	0
-------	-------	---

Notes: FOUND TO BE CORRECT

Engineer Name:

KS

Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
20700	20700	0
30700	30700	0
44000	44000	0

Repeatability


44000	44000	0
-------	-------	---

Traceable Weights Used: W24017

Trading Standards Maximum Tolerance: 60 +/- 60 kg

Calibration method and equipment used traceable to national standards.

QMF17 Issue 5 14/07/2020



Allweighs Weighing Services
 Westend Grange
 West End Road
 Epworth
 Doncaster
 DN9 1LA
 Tel: 01427 875 867
 Web: www.allweighs.co.uk



Certificate No: KCS:

202401045

Location:

Site Weighbridge out

Manufacturer:

Dini Argeo DFWL

Serial/Plant No:

0702402559

Capacity:

50000

kg

Increment:

20

kg

Frequency:

12

months

Test Date:

21 August 2024

Next due:

August 2025

Calibration Certificate of Accuracy

Business Name:

Transwaste Ltd

Address:

Melton Industrial Estate
 Gibson Lane South
 Melton
 Hull
 HU14 3HH

Contact Name:

Graham Wildridge

Off-centre Loading:

Load point:	Reading:	Deviation:	Load point:	Reading:	Deviation:
1	6700	0	5	6700	
2	6700	0	6	6700	
3	6700	0			
4	6700	0			

Four loadcell
 Six loadcell

Single loadcell
 Eight loadcell

As Found

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
20700	20700	0
30700	30700	0
44000	44000	0

Repeatability

44000	44000	0
-------	-------	---

Notes:

FOUND TO BE CORRECT

Engineer Name:

KS

Definitive

Measurement Points:	Reading:	Deviation:
0	0	0
8000	8000	0
16000	16000	0
20700	20700	0
30700	30700	0
44000	44000	0

Repeatability

44000	44000	0
-------	-------	---

Traceable Weights Used:

W24017

Trading Standards Maximum Tolerance:

60

+/-

60

kg

Calibration method and equipment used traceable to national standards.

QMF17 Issue 5 14/07/2020